

## CERTIFICATE OF ANALYSIS

### hCov-HKU1-E Pseudovirus Standard Reference

#### Product information

Name	hCov-HKU1-E Pseudovirus Standard Reference
Catalog No.	CBV30016
Lot No.	CGD2021030402
Format	Virus
Gene	Gene E
Sequence	249bp: ATGGTTGATTTATTTTCAATGATACTGCTTGGTACATAGGACAGATTTTAGTT TTAGTTTTATTTTGTCTTATTTCTTTAATCTTGTGTTGCTTTTTAGCAACTA TTAAGCTTTGTATGCAACTTTGTGGTTTTGTAATTTCTTTATTATTTACCTTC GGCTTACGTTTATAAAAGAGGTATGCAGTTGTATAAGTCTTATAGTGAACAAG TTATACCACCCACTTCAGATTATTTAATCTAA
NCBI No.	KF686346
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	03.04.2021
Expiry	12 months from the date of manufacture


#### Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*1vial	1ml/vial*1vial	√
Copy Number	>1x10 <sup>7</sup> copies/ml	2.24*10 <sup>7</sup> copies/ml	√
DNA Residue	<2%	1.58%	√

#### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*