

CERTIFICATE OF ANALYSIS

2019-nCoV-Panel Pseudovirus Standard Reference

Product information

Name	2019-nCoV-Panel Pseudovirus Standard Reference
Catalog No.	CBV30023
Lot No.	CGD2020090311
Format	Virus
Gene	Gene N, Gene E, ORF1ab/RdRP
Sequence	<p>Gene N (1260bp):</p> <pre> ATGTCTGATAATGGACCCCAAATCAGCGAAATGCACCCCGCA TTACGTTTGGTGGACCCTCAGATTCAACTGGCAGTAACCAGAA TGGAGAACGCAGTGGGGCGCGATCAAACAACGTCCGGCCCCA AGGTTTACCCAATAATACTGCGTCTTGGTTCACCGCTCTCACTC AACATGGCAAGGAAGACCTTAAATTCCCTCGAGGACAAGGCGT TCCAATTAACACCAATAGCAGTCCAGATGACCAAATTGGCTAC TACCGAAGAGCTACCAGACGAATTCGTGGTGGTGACGGTAAAA TGAAAGATCTCAGTCCAAGATGGTATTTCTACTACCTAGGAAC TGGGCCAGAAGCTGGACTTCCCTATGGTGCTAACAAAGACGGC ATCATATGGGTTGCAACTGAGGGAGCCTTGAATACACCAAAG ATCACATTGGCACCCGCAATCCTGCTAACAATGCTGCAATCGT GCTACAACCTCCTCAAGGAACAACATTGCCAAAAGGCTTCTAC GCAGAAGGGAGCAGAGGGCGGCAGTCAAGCCTCTTCTCGTTTCT CATCACGTAGTCGCAACAGTTCAAGAAATTCAACTCCAGGCAG CAGTAGGGGAACTTCTCCTGCTAGAATGGCTGGCAATGGCGGT GATGCTGCTCTTGCTTTGCTGCTGCTTGACAGATTGAACCAGCT TGAGAGCAAAATGTCTGGTAAAGGCCAACAAACAAGGCCA AACTGTCACTAAGAAATCTGCTGCTGAGGCTTCTAAGAAGCCT CGGCAAAAACGTACTGCCACTAAAGCATACAATGTAACACAA GCTTTCGGCAGACGTGGTCCAGAACAACCCAAGGAAATTTG GGGACCAGGAACTAATCAGACAAGGAACTGATTACAAACATT GGCCGCAAATTGCACAATTTGCCCCAGCGCTTCAGCGTTCTTC GGAATGTCGCGCATTGGCATGGAAGTCACACCTTCGGGAACGT GGTTGACCTACACAGGTGCCATCAAATTGGATGACAAAGATCC AAATTTCAAAGATCAAGTCATTTTGTGCTGAATAAGCATATTGAC GCATACAAAACATTCCCACCAACAGAGCCTAAAAAGGACAAA AAGAAGAAGGCTGATGAACTCAAGCCTTACCGCAGAGACAG AAGAAACAGCAAACCTGTGACTCTTCTTCCCTGCTGCAGATTGG ATGATTTCTCAAACAATTGCAACAATCCATGAGCAGTGCTGA CTCAACTCAGGCCTAA </pre> <p>Gene E (228bp):</p> <pre> ATGTA CTCA TTTCGTTTCGGAAGAGACAGGTACGTTAATAGTTA ATAGCGTACTTCTTTTCTTGCTTTCGTGGTATTCTTGCTAGTTA CACTAGCCATCCTTACTGCGCTTCGATTGTGTGCGTACTGCTGC </pre>

AATATTGTAAACGTGAGTCTTGTAACCTTCTTTTTACGTTT
CTCTCGTGTAAAAATCTGAATTCTTCTAGAGTTCCTGATCTC
TGGTCTAA

ORF1ab/RdRP (2995bp)

GTTGTCTGTAAGTCCGTTGCCACATAGATCATCCAAATCCTAAA
GGATTTTGTGACTTAAAAGGTAAGTATGTACAAATACCTACAA
CTTGTGCTAATGACCCTGTGGGTTTTACACTTAAAAACACAGTC
TGTACCGTCTGCGGTATGTGGAAAGGTTATGGCTGTAGTTGTG
ATCAACTCCGCGAACCATGCTTCAGTCAGCTGATGCACAATC
GTTTTTAAACGGGTTTGCAGGTGTAAGTGCAGCCCGTCTTACACC
GTGCGGCACAGGCACTAGTACTGATGTCGTATACAGGGCTTTT
GACATCTACAATGATAAAGTAGCTGGTTTTGCTAAATTCCTAA
AACTAATTGTTGTCGCTTCCAAGAAAAGGACGAAGATGACAA
TTTAATTGATTCTTACTTTGTAGTTAAGAGACACACTTTCTCTA
ACTACCAACATGAAGAAACAATTTATAATTTACTTAAGGATTG
TCCAGCTGTTGCTAAACATGACTTCTTTAAGTTTAGAATAGACG
GTGACATGGTACCACATATATCACGTCAACGTCTTACTAAATA
CACAATGGCAGACCTCGTCTATGCTTTAAGGCATTTTGATGAA
GGTAATTGTGACACATTAAGAAATACTTGTACATACAATT
GTTGTGATGATGATTATTTCAATAAAAAGGACTGGTATGATTTT
GTAGAAAACCCAGATATATTACGCGTATACGCCAACTTAGGTG
AACGTGTACGCCAAGCTTTGTTAAAAACAGTACAATTCTGTGA
TGCCATGCGAAATGCTGGTATTGTTGGTGTACTGACATTAGAT
AATCAAGATCTCAATGGTAACTGGTATGATTTCCGGTGATTTTCAT
ACAAACCACGCCAGGTAGTGGAGTTCCTGTTGTAGATTCTTATT
ATTCATTGTTAATGCCTATATTAACCTTGACCAGGGCTTTAACT
GCAGAGTCACATGTTGACACTGACTTAACAAAGCCTTACATTA
AGTGGGATTTGTTAAAATATGACTTCACGGAAGAGAGGTTAAA
ACTCTTTGACCGTTATTTTAAATATTGGGATCAGACATAACCACC
CAAATTGTGTTAACTGTTTGGATGACAGATGCATTCTGCATTGT
GCAAACCTTAATGTTTTATTCTCTACAGTGTCCACCTACAAG
TTTTGGACCACTAGTGAGAAAAATTTTGTGATGGTGTTCAT
TTGTAGTTTCAACTGGATACCCTTCAGAGAGCTAGGTGTTGTA
CATAATCAGGATGTAACTTACATAGCTCTAGACTTAGTTTTAA
GGAATTACTTGTGTATGCTGCTGACCCTGCTATGCACGCTGCTT
CTGGTAATCTATTACTAGATAAACGCACTACGTGCTTTTCAGTA
GCTGCACTTACTAACAATGTTGCTTTTCAAAGTCAAACCCCG
TAATTTTAAACAAGACTTCTATGACTTTGCTGTGTCTAAGGGTT
TCTTTAAGGAAGGAAGTTCTGTTGAATTAACACTTCTTCTTT
GCTCAGGATGGTAATGCTGCTATCAGCGATTATGACTACTATC
GTTATAATCTACCAACAATGTGTGATATCAGACAACACTACTATT
GTAGTTGAAGTTGTTGATAAGTACTTTGATTGTTACGATGGTGG
CTGTATTAATGCTAACCAAGTCATCGTCAACAACCTAGACAAA
TCAGCTGGTTTTCCATTTAATAAATGGGGTAAGGCTAGACTTTA
TTATGATTCAATGAGTTATGAGGATCAAGATGCACTTTTCGCAT
ATACAAAACGTAATGTCATCCCTACTATAACTCAAATGAATCT
TAAGTATGCCATTAGTGCAAAGAATAGAGCTCGCACCGTAGCT
GGTGTCTCTATCTGTAGTACTATGACCAATAGACAGTTTCATCA

AAAATTATTGAAATCAATAGCCGCGCCACTAGAGGAGCTACTGTA
 GTAATTGGAACAAGCAAATTCTATGGTGGTTGGCACAACATGT
 TAAAACTGTTTATAGTGATGTAGAAAACCCTCACCTTATGGG
 TTGGGATTATCCTAAATGTGATAGAGCCATGCCTAACATGCTT
 AGAATTATGGCCTCACTTGTCTTGTCTCGCAAACATACAACGTG
 TTGTAGCTTGTACACCGTTTCTATAGATTAGCTAATGAGTGTG
 CTC AAGTATTGAGTGAAATGGTCATGTGTGGCGGTTCACTATA
 TGTTAAACCAGGTGGAACCTCATCAGGAGATGCCACAACCTGCT
 TATGCTAATAGTGTTTTTAACATTTGTCAAGCTGTCACGGCCAA
 TGTTAATGCACTTTTATCTACTGATGGTAACAAAATTGCCGATA
 AGTATGTCCGCAATTTACAACACAGACTTTATGAGTGTCTCTAT
 AGAAATAGAGATGTTGACACAGACTTTGTGAATGAGTTTTACG
 CATATTTGCGTAAACATTTCTCAATGATGATACTCTCTGACGAT
 GCTGTTGTGTGTTTCAATAGCACTTATGCATCTCAAGGTCTAGT
 GGCTAGCATAAAGAAGCTTTAAGTCAGTTCTTTATTATCAAAC
 AATGTTTTTATGTCTGAAGCAAAATGTTGGACTGAGACTGACC
 TTAATAAGGACCTCATGAATTTTGTCTCAACATACAATGCTA
 GTTAAACAGGGTGATGATTATGTGTACCTTCCTTACCCAGATCC
 ATCAAGAATCCTAGGGGCGGCTGTTTTGTAGATGATATCGTA
 AAAACAGATGGTACACTTATGATTGAACGGTTCGTGTCTTTAG
 CTATAGATGCTTACCACTTACTAAACATCCTAATCAGGAGTAT
 GCTGATGTCTTTCATTTGTA CTTACAATACATAAGAAAGCTACA
 TGATGAGTTAACAGGACACATGTTAGACATGTATTCTGTTATG
 CTTACTAATGATAACACTTCAAGGTATTGGGAACCTGAGTTTTA
 TGAGGCTATGTACACACCGCATAACAGTCTTACAG

NCBI No.	NC_045512.2
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	09.03.2020
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ml/vial*10vials	1ml/vial*10vials	√
N Copy Number	>1x10 ⁷ copies/ml (DdPCR)	1.84x10 ⁷ copies/ml	√
E Copy Number	>1x10 ⁷ copies/ml (DdPCR)	6.54x10 ⁷ copies/ml	√
ORF1ab/RdRP Copy Number	>1x10 ⁷ copies/ml (DdPCR)	2.48x10 ⁷ copies/ml	√
DNA Residue	<1% (DdPCR)	0.69% (ORF1ab/RdRP) 0.90% (Gene N) 0.98% (Gene E)	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****