

CERTIFICATE OF ANALYSIS

2019-nCov-M Pseudovirus Standard Reference

Product information

Name	2019-nCov-M Pseudovirus Standard Reference
Catalog No.	CBV30032
Lot No.	CGD2021040701
Format	Virus
Gene	Gene M
Sequence	669bp ATGGCAGATTCCAACGGTACTATTACCGTTGAAGAGCTTAAAAAGCTCCTTGAA CAATGGAACCTAGTAATAGGTTTCCTATTCCCTTACATGGATTTGTCTTCTACAA TTTGCCATAGCCAACAGGAATAGGTTTTTGTATATAATTAAGTTAATTTTCCCTC TGGCTGTTATGGCCAGTAACTTTAGCTTGTTTTGTGCTTGCTGCTGTTTACAGA AATAAATTGGATCACCAGGTGGAATTGCTATCGCAATGGCTTGCTTGTAGGCTTG ATGTGGCTCAGCTACTTCATTGCTTCTTTCAGACTGTTTGCGCGTACGCGTTCC ACTATTCTGACCAGACCGCTTCTAGAAAAGTGAACCTCGTAATCGGAGCTGTGATC CTTGCGTGACATCTTCGTATTGCTGGACACCATCTAGGACGCTGTGACATCAAG GACCTGCCATAAGAAATCACTGTTGCTACATCACGAACGCTTTCTTATTACAAA TTGGGAGCTTCGCAGCGGTAGCAGGTGACTCAGGTTTTGCTGCATACAGTCGC TACAGGATTGGCAACTATAAATTAACACAGACCATTCCAGTAGCAGTGACAAT ATTGCTTTGCTTGCTGACAGTAA
NCBI No.	NC_045512.2
Buffer	1xPBS
Storage Conditions	-80°C
Date of Manufacture	04.07.2021
Expiry	12 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	100ul/vial*1 vial	100ul/vial*1 vial	√
Copy Number	>1x10e7copies/ml	1.46x10e7copies/ml	√
DNA Residue	<2%	0.93%	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****