

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2021-11-29		
Product Description:	ASSAYED HUMAN MULTI-SERA		
Catalogue Number:	HE1532		
Batch Number:	1257UE		
Expiry Date:	2025-11-28		

Analyte	Lot No.	Lot No. Target 1257UE 30.21	Range		Result	Units
ALBUMIN	1257UE		N/A	N/A	30.23	g/L
AST	1257UE	158.00	N/A	N/A	160.52	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Date: 78M022. Hmarn. Approved By:

Quality Control Representative