

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-10-29

KIT DESCRIPTION: MATERNAL CONTROL LEVEL 3

CATALOGUE NO: MSS5026

LOT NUMBERS: 7946MS

EXPIRY DATE: 2023-10-28

SYSTEM OF ANALYSIS: Roche Cobas 8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein (KIU/I)	7946MS	77.0-116	96.3	100	N/A
				98.9	
Total Beta hCG (IU/I)	7946MS	18264 - 27396	22830	23291	N/A
				22469	

QC TESTED BY:		Davina Graham		
DATE:	24-May-21			
RESULT (PASS/FAIL):		PASS		
QC APPROVED BY:		a-Martin		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr. 5.P. Fitzgerald BSc, PhD DSc (Hons) FIAE Man. R. J. Fitzgerald, W. Fitzgerald OBE. R. P. Kelly BSc FCA, Randox Laboratories Emited is a company registered in Northern Ireland, Company Number NI15738, VAT Number. GB 353 0304-00.