

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2021-09-29			
KIT DESCRIPTION:	LIQUID SPECIFIC PROTEIN CONTROL LEVEL 3			
CATALOGUE NO:	PS2684			
BATCH NO:	635LPC			
EXPIRY DATE:	2023-09-28			

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
635LPC	CRP	68.2	54.6-81.8	67.1	mg/L
635LPC	TOTAL PROTEIN	113.0	90.4-136.0	109.3	g/L

APPROVED BY:

L. Pavari

**QUALITY CONTROL REPRESENTATIVE** 

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com